

TEST AND CONFORMITY CERTIFICATE

**ACCORDING TO
EN 10204 - 3.1**

NUOVA FIMA

Industrial Instrumentation For Pressure & Temperature

NUOVAFIMA S.p.a. - Cap.Soc. € 3.500.000 I.V.
Via C.Battisti, 59/61 - 28045 INVORIO (NO) Italy
Tel. +39 0322 253200 - Fax +39 0322 253232
www.nuovafima.com - email : info@nuovafima.com
Codice fiscale / Partita IVA 01719710038
Reg.Imp. NOVARA 10895/1999 - REA 193327

MESSRS./SPETT.


**AG-ME AB
JARVIHAANTIE 10 HALLY C
1800 KLAUKKALA-FINLANDIA (FIN)**

Date	Certificate	Nuova Fima Order	Purchase Order N.	Sheet
06-06-2006	0000002365	1571/OR/2006	10162/2 of 28-10-2005	1
Description			Q.ty	Test
1.10.3.A.C.—.AAD9.24M.4SC.MC1.MCF.C PRESS. GAUGE MGS10/3/A DN63, -1...24 BAR / R134A / R407C / R404A, 1/4" SAE MALE, 4° COLOUR DIAL, CUSTOMER'S LOGO, 1 COLOUR, CUSTOMER+NUOVA FIMA LOGO, TYPE "C" FLANGE,			350	A,B,C=+-1,60 % F.S., D= 30,3 BAR
1.10.3.A.C.—.AAD7.24M.4SC.MC1.MCF.C PRESS. GAUGE MGS10/3/A DN63, -1...9 BAR / R134A / R407C / R404A, 1/4" SAE MALE, 4° COLOUR DIAL, CUSTOMER'S LOGO, 1 COLOUR, CUSTOMER+NUOVA FIMA LOGO, TYPE "C" FLANGE,			350	A,B,C=+-1,60 % F.S., D= 11,5 BAR

TEST: A) VISUAL B) DIMENSIONAL C) ACCURACY D) OVER PRESSURE E) TEST PRESSURE
F) REGULATION MICRO G) RELIABILITY MICRO H) TEST TEMPERATURE I) MAX STATIC PRESS. L) DIFFERENTIAL RANGE
M) MAX TEMPERATURE N) OVER TEMPERATURE

WE HEREBY CERTIFY THAT THE SUPPLY IS IN CONFORMITY WITH SPECIFICATIONS, DRAWINGS AND TO THE ORDER WHO IS REFERRED. THE SUPPLY HAS BEEN POSITIVELY CHECKED AND TESTED IN ACCORDANCE WITH THE NUOVA FIMA S.P.A. SPECIFICATIONS AND PROCEDURES.

NOTES :

FINAL CONTROL DEPT	INSPECTOR	THIRD PART INSPECTION	QUALITY ASSURANCE
SIGNATURE 	SIGNATURE	SIGNATURE	SIGNATURE